

Optische IOL-Konstanten für den Zeiss IOL Master

IOL	nominal	Haigis	Hoffer Q	Holl.1	SRK/T
SBL-3 SBL-2	A = 118.0	a0 = 0.537 a1 = 0.333 a2 = 0.126	pACD = 5.22	sf = 1.47	A = 118.43
Tetraflex HD	A = 118.0	a0 = 1.02 a1 = 0.40 a2 = 0.10	pACD = 5.30	sf = 1.55	A = 118.57
NEU AURIUM Matrix 404	A = 118.3	a0 = 1.464 a1 = 0.4 a2 = 0.1	pACD = 4.96	sf = 1.46	A = 118.3
ARTIS [®] Toric	A = 119.3	a0 = 0.088 a1 = 0.233 a2 = 0.20	pACD = 6.095	sf = 2.295	A = 119.74
ARTIS [®] PL E	A = 119.3	a0 = 0.088 a1 = 0.233 a2 = 0.20	pACD = 6.095	sf = 2.295	A = 119.74
EAZ [®] -Y ●	A = 119.3	a0 = 1.77 a1 = 0.40 a2 = 0.10	pACD = 6.03	sf = 2.33	A = 119.7
NEU Softec HP1	A = 118.7		pACD = 5.37		A = 118.7
Softec HD Softec HDY ●	A = 118.0	a0 = 0.92 a1 = 0.40 a2 = 0.10	pACD = 5.22	sf = 1.47	A = 118.43
NEU Softec HD3	A = 118.0		pACD = 5.04	sf = 1.03	A = 118.13
Softec 1	A = 118.0	a0 = 0.92 a1 = 0.40 a2 = 0.10	pACD = 5.22	sf = 1.47	A = 118.43
CLARÉ [®]	A = 118.0	a0 = 1.59 a1 = 0.40 a2 = 0.10	pACD = 5.26	sf = 1.51	A = 118.5
SAL TORIC	A = 118.7	a0 = 1.32 a1 = 0.40 a2 = 0.10	pACD = 5.51	sf = 1.75	A = 118.9
SAL 302AC SAL 302A ●	A = 118.7	a0 = 1.32 a1 = 0.40 a2 = 0.10	pACD = 5.51	sf = 1.75	A = 118.9
SAL P302AC SAL P302A ●	A = 118.7	a0 = 1.32 a1 = 0.40 a2 = 0.10	pACD = 5.51	sf = 1.75	A = 118.9
SAL 300AC SAL 300A ●	A = 118.3	a0 = 1.26 a1 = 0.40 a2 = 0.10	pACD = 5.39	sf = 1.61	A = 118.7

Alle angegebenen A-Konstanten sind Mittelwerte. Wir empfehlen eigene Werte zu ermitteln, die gegebenenfalls abweichen können.