







A-Konstanten für die optische Biometrie

IOL	nominal	Haigis	Hoffer Q	Holl.1	SRK/T
NEU Alsafit Fourier  Alsaset Fourier Toric	A = 118.0	a0 = 0.183 a1 = 0.40 a2 = 0.10	pACD = 4.43	sf = 0.65	A = 117.3
NEU ARTIS [®] Symbiose	A = 119.3	a0 = 0.088 a1 = 0.233 a2 = 0.200	pACD = 6.095	sf = 2.295	A = 119.7
SBL-3 SBL-2	A = 118.0	a0 = 0.537 a1 = 0.333 a2 = 0.126	pACD = 5.22	sf = 1.47	A = 118.43
AURIUM Matrix 404	A = 118.3	a0 = 1.464 a1 = 0.4 a2 = 0.1	pACD = 4.96	sf = 1.46	A = 118.3
ARTIS [®] Toric	A = 119.3	a0 = 0.088 a1 = 0.233 a2 = 0.20	pACD = 6.095	sf = 2.295	A = 119.74
ARTIS [®] Monofokal	A = 119.3	a0 = 0.088 a1 = 0.233 a2 = 0.20	pACD = 6.095	sf = 2.295	A = 119.74
EAZ [®] -Y 	A = 119.3	a0 = 1.77 a1 = 0.40 a2 = 0.10	pACD = 6.03	sf = 2.33	A = 119.7
Softec HD Softec HDY 	A = 118.0	a0 = 0.92 a1 = 0.40 a2 = 0.10	pACD = 5.22	sf = 1.47	A = 118.43
Softec HD3	A = 118.0		pACD = 5.04	sf = 1.03	A = 118.13
Softec 1	A = 118.0	a0 = 0.92 a1 = 0.40 a2 = 0.10	pACD = 5.22	sf = 1.47	A = 118.43
CLARÉ [®]	A = 118.0	a0 = 1.59 a1 = 0.40 a2 = 0.10	pACD = 5.26	sf = 1.51	A = 118.5
SAL 302AC SAL 302A 	A = 118.7	a0 = 1.32 a1 = 0.40 a2 = 0.10	pACD = 5.51	sf = 1.75	A = 118.9
SAL P302AC SAL P302A 	A = 118.7	a0 = 1.32 a1 = 0.40 a2 = 0.10	pACD = 5.51	sf = 1.75	A = 118.9
SAL 300AC SAL 300A 	A = 118.3	a0 = 1.26 a1 = 0.40 a2 = 0.10	pACD = 5.39	sf = 1.61	A = 118.7

Alle angegebenen A-Konstanten sind Mittelwerte. Wir empfehlen eigene Werte zu ermitteln, die gegebenenfalls abweichen können.